

# EC DECLARATION OF CONFORMITY

We,

**Wagner & Guder Medical GmbH**  
**Dorfstraße 57, D- 99510 Hermstedt**

declare by ourself, that following product

Product description: DigiCam SL (Digital camera & Capture Software)  
Product name/ Model: DC-20, DC-30, DC3-30, DC-75, DC3-75, DC-105,  
DC-115, DC-120, DC3-120, DC-130, DC3-130,  
DC-500, DC-900BQ, DC3-900BQ, DC-SL2E,  
DC3D\_30, DC3D\_130, DC3D\_HU, DC3D\_HS

UMDNS- Code: camera system, ophthalmological (16-419)  
image data transmission system (17-960)

Classification: I

is produced and designed in accordance of the guideline 93/42/EEC and its transpositions in national laws which apply to it.

The technical documentation is kept at the address above.

Applicable harmonized standards have been applied.

Applied Standards: IEC 60601-1: 2013  
IEC 60601-1: 2005 + A1:2012  
DIN EN 60601-1-2: 2016  
ISO 14971: 2007  
ISO 10939  
ISO 15004

We declare that the EU Directive 2011/65/EU is fulfilled.



This declaration is valid for all devices after the date of manufacture and submitted by:

*S.G. Guder*

---

Stefan Guder, Dipl.-Ing.(FH)  
(safety manager)

Hermstedt, 02.01.2018